



Product Analysis Certificate

Art.Nr.: CL03.0216

Lot Nr.: 29.5892307

Buffer pH 7.00

pH 7.00 ($\pm 0.02/20^{\circ}\text{C}$) Phosphate

Measured value: pH = 7,00 - 20°C

Use: For laboratory use, pH measurements, traceable to SRM from NIST

Traceability: Reference standards acc. NIST SRM 189c, 188, 185i, 186lg, 186llg, 187e

Specifications:

pH = 7.13 @ 0°C
pH = 7.05 @ 10°C
pH = 7.00 @ 20°C
pH = 6.99 @ 25°C
pH = 6.98 @ 30°C
pH = 6.97 @ 40°C
pH = 6.96 @ 50°C
pH = 6.96 @ 60°C
pH = 6.97 @ 70°C
pH = 6.98 @ 80°C
pH = 7.00 @ 90°C

Chemist: Luis Bianchi

Date of release: 27 August 2021
Expires: Jul-2023
Version: 1

CHEM-LAB NV

Industriezone "De Arend" 2 B-8210 ZEDELGEM - BELGIUM

Tel.: +32 50 28 83 20 Fax.: +32 50 78 26 54 e-mail : info@chem-lab.be Web : www.chem-lab.be