Page: 1

Honeywell

CERTIFICATE OF ANALYSIS

Honeywell Specialty Chemicals Seelze GmbH

Wunstorferstrasse 40

Seelze, 30926

Telefon: +49 5137 999-0

lab.honeywell.com

Brand:

Honeywell Fluka

Product:

33668

Lot No: Production date:

N0870

Expiry Date:

28.Mar.2023

08.Sep.2026

Buffer solution pH 10,00 solution ready for use, violett coloured

Parameter	Specification	Units	Result	Units
pH (20 °C)	9.95-10.05		10.04	

Measured with a combined glass electrode. The calibrating of the pH value is carried out with standard buffer solutions (related to SRM by NIST) according to DIN 19266 at several reference points (according to DIN 19268).

Composition per litre:

11.76 g Citric acid-1-hydrate (C6H8O7*H2O)

2.57 g Sodium chloride (NaCl)

2.72 g Sodium hydroxide (NaOH)

Temperature-dependence of the pH values:

pH 10.24 at 0°C pH 9.75 at 50°C pH 10.10 at 10°C pH 9.68 at 60°C pH 9.90 at 30°C pH 9.82 at 40°C pH 9.49 at 90°C

The checking of the chemicals and of the test equipment used is in accordance with DIN ISO 9001.

QC Release Date: 22.May.2023

The minimum shelf life is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.

We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.

Particular properties of the products or the suitability for a particular area of application are not assured.

We guarantee a proper quality within our General Conditions of Sales.

Page: 2

Honeywell

Honeywell

Quality Management

This document was produced electronically and is valid without a signature